



LATEX STRONG 44-160





CHARACTERISTICS

- · Powdered latex glove
- · Smooth inside
- · Supple and pleasant to wear
- · Very easy to put on thanks to rolled edge
- · Suitable for the right and left hand (ambidextrous)
- · Suitable for contact with many foods
- \cdot Meets the EN 455 requirement for medical examination gloves
- · Length: 240 mm and thickness: 0.10 mm
- Weight: 6.5 grams (+/- 0.3 grams)
- AQL 1.5 (EN 374)
- · Gloves are for single use only
- · Packed per 100 pieces in a dispenser

Article number: 1.44.160.00

SUITABLE FOR ACTIVITIES IN E.G.

- · Food industry
- Agri
- Hospitality
- Maintenance
- · Cleanina
- Automotive

COLOUR

Natural

OTHER AVAILABLE COLOURS

Blue (art.no. 1.44.161.00)

SIZES

6/XS to 10/XL incl.

PACKAGING

- · 100 pieces per dispenser
- · 10 dispensers in a case

EN ISO 374-1:2016 +A1:2018/Type C








EN 455

C € 1639 EN 420:2003+A1:2009



PRODUCT INFORMATION

SIZE	ARTICLE NO.	EAN CODE 100 PIECES (DISPENSER)	EAN CODE 10 DISPENSERS (CASE)		
6/XS	1.44.160.06	8718249008245	8718249008252		
7/S	1.44.160.07	8718249008269	8718249008276		
8/M	1.44.160.08	8718249008283	8718249008290		
9/L	1.44.160.09	8718249008306	8718249008313		
10/XL	1.44.160.10	8718249008320	8718249008337		



CLARIFICATION OF PICTOGRAMS

Protection against chemicals and micro-organisms

EN ISO 374-1:2016/Type B					
EN ISO 374-1:2016 /Type A	Chemicals	EN 16523:2015 permeation level			
	40% Sodium Hydroxide (K)	6			

EN ISO	374-4:2019
CAS number (K) 1310-73-2	Class Inorganic base
Chemical 40% Sodium Hydroxide (K)	Average decrease % -33.0%

Resistance to bacteria & fungi and virus

Air leak test – Pass

EN ISO 374-5:2016					
EN ISO 374-2:2014	Resistance to bacteria & fungi	Virus resistance			
VIRUS	PASS	PASS			
EN ISO 374-2:2014					

Water leak test – Pass

EN ISO 374-1:2016+A1:2018 Permeation levels are based on the following lead times:

Performance level	1	2	3	4	5	6
Minimum breakthrough times (in min.)	>10	>30	>60	>120	>240	>480

STORAGE REQUIREMENTS

The gloves should be stored in a clean, cool and dry place and not kept compressed in the original packaging. Do not expose the gloves to direct sunlight. Make sure that the packaging and the gloves are not damaged when shipped.

TESTING INSTITUTE

These gloves are certified by: SATRA Technology Europe Ltd (Notified Body no. 2777), Bracetown Business Park, Clonee, Dublin D15 YN2P, Ireland.

DECLARATION OF CONFORMITY

For a copy of the declaration of conformity, please refer to the following web link: www.oxxa-safety.com/doc

RELATED ITEMS



LATEX-THIN 44-140ART. NO. 1.44.140.00



LATEX-STRONG 44-215 ART. NO. 1.44.215.00

YOUR SUPPLIER:
