



DEXTERITY




POWDER
FREE



CHEMICAL
RESISTANT

NITRI- STRONG 44-530



CHARACTERISTICS

- Unpowdered non sterile disposable nitrile glove
- Glove fully embossed
- Smooth inside
- Very easy to put on due to rolled hem
- Suitable for the right and left hand (ambidextrous)
- 100% latex free in favor of any latex allergies
- Silicone-free
- Length: 240mm and thickness: 0.13mm
- Weight: 6.5 grams (+/- 0.3 grams)
- AQL 1.5 (EN 374)
- Suitable for contact with many foodstuffs
- Meets EN 455 requirement for medical research gloves
- Packed per 100 pieces in a dispenser

Article number: 1.44.530.00

SUITABLE FOR ACTIVITIES IN E.G.

- Food industry
- Agri
- catering industry
- Maintenance
- Chemical sector
- Automotive

COLOUR

Blue

SIZES

7/S to 11/XL

PACKAGING

- 100 pieces per dispenser
- 10 dispensers per outer carton

EN ISO 374-1:2016
+A1:2018/Type B



KPT

EN ISO 374-5:
2016




VIRUS



CE 2777
EN 420:2003+A1:2009

PRODUCT INFORMATION

SIZE	ARTICLE NO.	EAN-CODE 100 PIECES (DISPENSER)	EAN-CODE 10 DISPENSERS (OUTER CARTON)
7/S	1.44.530.07	8718249008689	8718249065613
8/M	1.44.530.08	8718249008702	8718249065620
9/L	1.44.530.09	8718249008726	8718249065637
10/XL	1.44.530.10	8718249008740	8718249065644
11/XXL	1.44.530.11	8718249008764	8718249065651



EXPLANATION OF ICONS

Protection against chemicals and microorganisms

EN ISO 374-1:2016/Type B		
EN ISO 374-1:2016 /Type A	Chemical	EN 16523:2015 permeation level
	40% Sodium Hydroxide (K) 30% Hydrogen peroxide (P) 37% Formaldehyde (T)	6 4 6

EN ISO 374-4:2019	
Cas number (K) 1310-73-2 (P) 7722-84-1 (T) 50-00-0	Class Inorganic base Peroxide Aldehyde
Chemical	Gemiddelde degradatie %
40% Sodium Hydroxide (K) 30% Hydrogen peroxide (P) 37% Formaldehyde (T)	-3.9% 8.1% 2.9%

Resistance to Bacteria & Fungi and Virus

EN ISO 374-5:2016		
EN ISO 374-2:2014	Resistance to bacteria and fungi	Resistance to Virus
	PASS	PASS

EN ISO 374-2:2014	
Air leak test - Pass	Water leak test - Pass

EN ISO 374-1:2016 Permeation levels are based on the following lead times:

Prestation level	1	2	3	4	5	6
Minimum lead times (in min.)	>10	>30	>60	>120	>240	>480

STORAGE INSTRUCTIONS

The gloves should be stored in a clean, cool and dry place and not compressed in the original packaging. Do not expose the gloves to direct sunlight. Please ensure that the packaging and gloves are not damaged during shipping.

TEST INSTITUTE

These gloves are certified by: SATRA Technology Europe Ltd (Notified Body no. 2777), Bracetown Business Park, Cloness, Dublin D15 YN2P, Ireland.

DECLARATION OF CONFORMITY

For a copy of the declaration of conformity, we refer you to the following link: www.oxxa.work/doc

YOUR SUPPLIER